



Annex (1)

Updated on: 09-01-2024

To the Accreditation Certificate No. **JAS Cal. - 010** Dated **26-12-2022**

For The laboratory at Validation for Information Technology (Technival) / Amman

Scope of Accreditation

In the field of Temperature, Humidity, and Pressure Calibration (Permanent and onsite)

Measurand	Measuring Range	Calibration and measurement Capability (CMC) ^a	Calibration Methods/ Standards/ Remarks
Thermocouples (Permanent and onsite)			
Thermocouples	≥ -40 to 0 °C	0.19 °C	SOP100 Tech QP-004 [Issue No. (1); Issue Date: 15/03/2021, Revision No. (5); Revision Date:18/10/2022] according to EURAMET cg-8 Version 3.1 (02/2020) - Calibration of Thermocouples
	> 0 to 100 °C	0.23 °C	
	> 100 to 400 °C	1.4 °C	
	> 400 to 650 °C	5.5 °C	
Temperature and Humidity Sensors (Permanent and onsite)			
Temperature and Humidity Sensors	≥ 0 to 60 °C	0.60 °C	SOP101 Tech QP-004 [Issue No. (1); Issue Date: 15/03/2021, Revision No. (5); Revision Date:18/10/2022] according to EURAMET cg-8Version 3.1 (02/2020) - Calibration of Temperature Data Loggers



Annex (1)

Updated on: 09-01-2024

To the Accreditation Certificate No. **JAS Cal. - 010** Dated **26-12-2022**

For The laboratory at Validation for Information Technology (Technival) / Amman

Scope of Accreditation

In the field of Temperature, Humidity, and Pressure Calibration (Permanent and onsite)

Measurand	Measuring Range	Calibration and measurement Capability (CMC) ^a	Calibration Methods/ Standards/ Remarks
	≥ 5 % to 95 %	3.9 % RH	SOP102 Tech QP-004 [Issue No. (1); Issue Date: 15/03/2021, Revision No. (5); Revision Date: 18/10/2022] according to DKD-R 5-8 Version: 10/2019 Revision: 0 - Calibration of Humidity Sensors and Data Loggers
Pressure Gauges (Permanent and onsite)			
Pressure Gauge-Pneumatic	-0.8 Bar to 20 Bar	0.02 Bar	SOP103 Tech QP-004 [Issue No. (1); Issue Date: 15/03/2021, Revision No. (5); Revision Date: 18/10/2022] according to DKD-R 6-1 Version 03/2014 Revision: 3 - Calibration of Pressure Indicators, Sequence "C" Only.
Pressure Gauge-Hydraulic	20 bar to 100 bar	0.50 Bar	SOP103 Tech QP-004 [Issue No. (1); Issue Date: 15/03/2021, Revision No. (5); Revision Date: 18/10/2022] according to DKD-R 6-1 Version 03/2014 Revision: 3 - Calibration of Pressure Indicators, Sequence "C" Only.



Annex (1)

Updated on: 09-01-2024

To the Accreditation Certificate No. **JAS Cal. - 010** Dated **26-12-2022**

For The laboratory at Validation for Information Technology (Technival) / Amman

Scope of Accreditation

In the field of Temperature, Humidity, and Pressure Calibration (Permanent and onsite)

a) The reported CMCs are expressed at approximately the 95 % level of confidence, using a coverage factor of $k = 2$. The actual measurement uncertainty of a specific calibration performed by the laboratory may be greater than the CMC due to the behavior of the customer's device and to influences from the circumstances of the specific calibration.

List of employees in the laboratory who are technically responsible for issuing the calibration certificates in the scope of accreditation:

1. Malek Shamlawe: General Manager
2. Afnan Dawoud: Quality Manager
3. Issa Jamal: Executive Manager
4. Mohammad Rawashdeh: Calibration and Qualification Supervisor
5. Aws AL-Batainah: Calibration and Qualification Senior
6. Hasan Rawashdeh: Calibration and Qualification Employee
7. Reziq Rawashdeh: Calibration and Qualification Senior