



THE HASHEMITE KINGDOM OF
JORDAN
Accreditation Unit



Annex (1)

Updated on 09/11/2025

To the Accreditation Certificate No. JAS Cal. - 013 Dated 30-09-2025

**For the Laboratory of Jordanian Elite Company for the Repair and Maintenance of
Measuring and Testing Equipment / Amman**

Scope of Accreditation

Onsite Calibration of Non-Automatic Weighing Instruments

Measurand	Measuring Range	Calibration and measurement Capability (CMC) ^a	Calibration Methods/ Standards/ Remarks
Mass			
Non-Automatic Weighing Instruments - class III	1 kg to $\leq$ 2000 kg	6 g	OIML R76:2006 - Non- Automatic Weighing Instrument
Weighbridges - Class III and IV	1 ton to $\leq$ 10 ton	3 kg	
	10 ton to $\leq$ 20 ton	6 kg	
	20 ton to $\leq$ 32 ton	12 kg	

a) The reported CMCs are expressed at approximately the 95 % level of confidence, using a coverage factor of $k = 2$. The actual measurement uncertainty of a specific calibration performed by the laboratory may be greater than the CMC due to the behavior of the customer's device and to influences from the circumstances of the specific calibration.

List of employees in the laboratory who are technically responsible for issuing the calibration certificates in the scope of accreditation:

1. General Manager: Dr. Zakariya Hamad Inad
2. The head of the calibration department: Eng. Anas Yasser Alzaareer
3. Calibration technician: Mr. Amr Gamal Salah

qf071-79, rev. b