



Accreditation Unit

Annex (1)

To The Accreditation Certificate No. **JAS Med.- 010** Dated **09/02/2026**for Laboratory of ACDIMA Center for Bioequivalence and Pharmaceutical Studies (ACDIMA
Biocenter) / Amman

Scope of Accreditation

In the Fields of Clinical Chemistry and Hematology

SOP	Main Profile (Sub-Profile)	Parameter/ Analyte	Sample type	Test method	Instrument
Clinical Chemistry (Routine)					
SP10-008, T.01	Clinical Chemistry (Routine)	ALT	Serum	IFCC UV Without pyridoxal phosphate method	C311, C111
SP10-008, T.01	Clinical Chemistry (Routine)	AST	Serum	IFCC UV Without pyridoxal phosphate method	C311, C111
SP10-008, T.01	Clinical Chemistry (Routine)	ALP	Serum	IFCC Colorimetric method	C311, C111
SP10-008, T.01	Clinical Chemistry (Routine)	Creatinine	Serum	Enzymatic method	C311, C111
SP10-008, T.01	Clinical Chemistry (Routine)	Urea	Serum	Kinetic test with urease and glutamate dehydrogenas method	C311, C111
SP10-008, T.01	Clinical Chemistry (Routine)	Total Protein	Serum	Colorimetric Biuret method	C311, C111
SP10-008, T.01	Clinical Chemistry (Routine)	Total Bilirubin	Serum	Colorimetric diazo method	C311, C111
SP10-008, T.01	Clinical Chemistry (Routine)	Glucose	Serum	Hexokinase method	C311, C111
SP10-008, T.01	Clinical Chemistry (Routine)	Sodium	Serum	Indirect Ion-Selective Electrode (ISE)	C311, AVL9180

**Accreditation Unit****Annex (1)****To The Accreditation Certificate No. JAS Med.- 010 Dated 09/02/2026****for Laboratory of ACDIMA Center for Bioequivalence and Pharmaceutical Studies (ACDIMA Biocenter) / Amman****Scope of Accreditation****In the Fields of Clinical Chemistry and Hematology**

SOP	Main Profile (Sub-Profile)	Parameter/Analyte	Sample type	Test method	Instrument
SP10-008, T.01	Clinical Chemistry (Routine)	Potassium	Serum	Indirect Ion-Selective Electrode (ISE)	C311, AVL9180
Hematology (Routine)					
SP10-008, T.01	Hematology (Routine)	Hemoglobin	EDTA Whole blood	Non cyanide method	SysmexXP300
SP10-008, T.01	Hematology (Routine)	Hematocrit	EDTA Whole blood	DC detection method	SysmexXP300
SP10-008, T.01	Hematology (Routine)	WBC	EDTA Whole blood	DC detection method	SysmexXP300
SP10-008, T.01	Hematology (Routine)	RBC	EDTA Whole blood	DC detection method	SysmexXP300
SP10-008, T.01	Hematology (Routine)	Platelet	EDTA Whole blood	DC detection method	SysmexXP300
SP10-008, T.01	Hematology (Routine)	MCV	EDTA Whole blood	Calculated	SysmexXP300
SP10-008, T.01	Hematology (Routine)	MCH	EDTA Whole blood	Calculated	SysmexXP300
SP10-008, T.01	Hematology (Routine)	MCHC	EDTA Whole blood	Calculated	SysmexXP300

List of employees in the laboratory who are technically responsible for issuing the test reports in the scope of accreditation:

1. Medical Lab Director: Dr. Concepsion Esbinar Basue
2. Medical Lab Director Deputy: Dr. Usamah Sayed